

FAST MULTI-LEVEL RECONSTRUCTION OF BIOMEDICAL IMAGES USING WAVELET SPARSITY CONSTRAINTS**Michael Unser¹, Cédric Vonesch¹, Matthieu Guerquin-Kern¹ and Dimitri Van De Ville^{1,2}**¹Biomedical Imaging Group, EPFL, Switzerland²University Hospital Geneva, Switzerland

Wavelet-domain ℓ_1 -regularization is a powerful approach for solving inverse problems. In their 2004 landmark paper, Daubechies et al. proved that one could solve such linear inverse problems by means of a “thresholded Landweber” (TL) algorithm [1]. While this iterative procedure is simple to implement, it is known to converge slowly. Here, we present a multilevel version of the algorithm that is inspired from the multigrid techniques used for solving PDEs, but with one important difference : instead of cycling through coarser versions of the problem (REDUCE part of multigrid), the multilevel algorithm cycles through the successive wavelet subspaces. The method works with arbitrary wavelet representations ; it typically yields a 10-fold speed increase over the standard TL algorithm, while providing the same restoration quality. We illustrate the applicability of the method to three biomedical image reconstruction problems : the deconvolution of 3D fluorescence micrographs [2], the global reconstruction of dynamic PET from time measurements [3], and the reconstruction of magnetic resonance images from arbitrary (non-uniform) k -space trajectories. We present experimental results with real data sets in all three cases.

References

- [1] I. Daubechies, M. Defrise, and C. De Mol, An iterative thresholding algorithm for linear inverse problems with a sparsity constraint, *Comm. Pure Appl. Math.*, **57**, (2004), 1413–1457.
- [2] C. Vonesch, M. Unser, A Fast Multilevel Algorithm for Wavelet-Regularized Image Restoration, *IEEE Trans. Image Processing*, **18(3)**, (2009), 509–523.
- [3] J. Verhaeghe, D. Van De Ville, I. Khalidov, Y. D’Asseler, I. Lemahieu, M. Unser, Dynamic PET Reconstruction Using Wavelet Regularization with Adapted Basis Functions, *IEEE Trans. Medical Imaging*, **27(7)**, (2008), 943–959.